



CERTIFICATE OF CALIBRATION



Glanford Electronics Ltd.

Glanford House,
Exmoor Avenue
Skippingdale Business Park,
Scunthorpe
North Lincolnshire
DN15 8NJ

Telephone: (01724) 289600 Fax: (01724) 289008
Email: info@glanfordelectronics.com www.glanfordelectronics.com

CALIBRATION: 2900

Issued to: Trace Measurement Limited
Essex

Instrument Certified: Druck DPI 700 Digital Pressure Indicator

Serial No: 16660/96-2

Date in: 16-Nov-11

Cal Date: 22-Nov-11

Date Due: 22-Nov-12

All measurements are traceable to UK National Standards / a UKAS Standard

(GEL201) Druck DPI740 Pressure Calibrator S/N:74001459, (GEL205)
Pressurements Air Dead Weight Tester S/N:4848-87

Observations/Corrections

Repaired and Adjusted

New Battery Fitted

Calibrated at 23 °C 43 %RH

Certified that, unless otherwise stated, the equipment conforms in all respects to manufacturers published specifications at the points measured
Calibrations performed in accordance with our registration to BS EN ISO9001:2008
and the general requirements of BS EN ISO 10012:2003

certifying engineer

S. Lawrence
J. Taylor ✓
M. Gibson
C. Cowling
G. Cowling